

Certificate

on the antimicrobial properties of materials
The product

MEDI grip

of the company

HEINRICH KIPP WERK KG
Heubergstraße 2, 72172 Sulz am Neckar, Deutschland

has been successfully tested against
Staphylococcus epidermidis DSM 18857
Pseudomonas aeruginosa DSM 939 / ATCC 15442

as antimicrobial
(≥ 99.9% germ reductionreduktion)

according to ASTM E 2180 (modified)

For details, see Test Reports No. 2202.1 / 150922-683-2180-01; sample: 6832109154077;
2407 / 160511-683-2180-01; sample: 6832704164483

The bearer of this certificate is entitled to affix the following seal to
parts or surfaces made from the tested material



Nuremberg, Germany, May 18th, 2016


Managing Director


Laboratory Director

QualityLabs BT GmbH, Neumeyerstraße 46a, 90411 Nuremberg, Germany

The issuer of this certificate is an accredited and recognized testing laboratory for testing
antimicrobial and adhesive surface properties of materials used in medical devices.
Register number: D-PL-13335-01-00 / ZLG-AP-231.10.72
Competency certification according to directives 93/42/EWG, 90/385/EWG and DIN EN ISO/IEC
17025.

CERTIFIED according to ASTM E 2180